



Evaluation of Wearable Technologies in Physiotherapy and Rehabilitation

 Okan Şahin,^{1,2}  Mustafa Oğuz Kethüdaoğlu,^{2,3}  Mehmet Kaan Altunok,^{2,4}  Aygül Köseoğlu Kurt,²
 Suzan Aydın,^{2,5}  Harun Dere,²  Ali Veyssel Özden²

¹Department of Physiotherapy and Rehabilitation, İstanbul Health and Technology University Faculty of Health Sciences, İstanbul, Türkiye

²Department of Physiotherapy and Rehabilitation, Bahçeşehir University, Graduate Education Institute Doctorate Program, İstanbul, Türkiye

³Department of Physiotherapy and Rehabilitation, Cumhuriyet University Faculty of Health Sciences, Sivas, Türkiye

⁴Department of Physiotherapy and Rehabilitation, Selçuk University Faculty of Health Sciences, Konya, Türkiye

⁵Department of Physiotherapy and Rehabilitation, İstanbul Gelişim University Faculty of Health Sciences, İstanbul, Türkiye

Abstract

Wearable technologies are defined as electronic devices which can be carried on clothes and accessories and provide psychological monitoring and biological feedback as well. Rings, bracelets, watches, jewelry such as lenses or spectacles, headbands, e-textile goods, and hearing aids are all types of wearable devices. In this study, “wearable technologies,” “physiotherapy,” “physical activity,” “sleep,” and “smart textiles” keywords have been searched in PubMed, Google Scholar, Cochrane, and Scopus databases between 2001 and 2023. Wearable devices used in central nervous system diseases, athletes, sleep, physical activity, and weight control as well as smart textiles used in rehabilitation have been included in our study. Wearable technology makes it possible to expand the database of health-related information in addition to helping to monitor, improve, and maintain health. These technologies are commonly used in many diseases’ diagnoses such as epilepsy and sleep disorders, in the follow-up of parameters related to physical activity and athlete’s health, treatment, and rehabilitation processes of central nervous system diseases like Alzheimer’s, and Parkinson’s. In addition, smart textile products are used for sensory processing problems of individuals with autism spectrum disorder and figure out the risk of diabetic foot ulcers. By examining wearable technologies used for different purposes in physiotherapy and rehabilitation, this study contributes to the use of these technologies in new scientific studies.

Keywords: Physical activity, physiotherapy and rehabilitation, sleep, smart textiles, wearable technologies.

Cite This Article: Şahin O, Kethüdaoğlu MO, Altunok MK, Köseoğlu Kurt A, Aydın S, Dere H, Özden AV. Evaluation of Wearable Technologies in Physiotherapy and Rehabilitation. BAU Health Innov 2023;1(2):93–100.

Wearable technologies are defined as electronic devices which can be carried on clothes and accessories and provide psychological monitoring and biological feedback as well. Rings, bracelets, watches, jewelry such as lenses or spectacles, headbands, e-textile goods, and hearing aids are all types of wearable devices.^[1]

Wearable devices are considered one of the important applications of devices that can send and receive data

over the internet, now called objects of the internet. Due to the widespread use of the Internet of Things, users’ opportunities to take control of their health and make healthy life choices have increased thanks to devices that can be connected through the internet.^[2]

Wearable technologies can be used for different purposes, including health monitoring, management of chronic diseases, diagnosis and treatment of diseases,

Address for correspondence: Okan Şahin, MD. İstanbul Sağlık ve Teknoloji Üniversitesi Sağlık Bilimleri Fakültesi, Fizyoterapi ve Rehabilitasyon Bölümü, İstanbul, Türkiye

Phone: +90 545 279 66 08 **E-mail:** okan.sahin@istun.edu.tr

Submitted: September 29, 2023 **Revised:** November 15, 2023 **Accepted:** November 24, 2023 **Available Online:** May 08, 2024

BAU Health and Innovation - Available online at www.bauhealth.org

OPEN ACCESS This work is licensed under a Creative Commons Attribution-NonCommercial 4.0 International License.



and rehabilitation.^[3] In addition to allowing the user to monitor and improve their health status, these devices also contribute to the significant development of medical technology thanks to the sensors they contain.

^[4] Wearable devices can contain a wide variety of sensors such as accelerometers that measure acceleration, electrocardiograms that measure the electrical activity of the heart, galvanic skin receptor sensors that measure stress level and emotional state, glucose monitors that measure blood glucose levels, or photoplethysmography that measures blood volume.^[5]

The use of wearable devices is widespread in the diagnosis, follow-up, treatment, and rehabilitation processes of many diseases. These devices provide significant benefits to both the wearable technology user and the healthcare professionals who monitor the user during the treatment process of musculoskeletal system problems or cognitive diseases.^[3] It is estimated that the use of wearable devices designed to monitor physical activity and weight control will increase more than fivefold every half year.^[6]

Wearable technologies are also used to detect sleep disorders. Sleep plays an important role in people's daily activities, and sleep disorders seriously affect the quality of life. Many studies have shown the relationship between poor sleep quality and many negative effects on our health, including but not limited to obesity, diabetes, heart disease, hypertension, mood disorders, weakened immune system, and increased risk of death.

^[7,8] To treat sleep disorders, it is very important to first detect, record, monitor, and analyze the disturbed sleep pattern. In the literature, polysomnography (PSG) is accepted as the gold standard in traditional sleep monitoring and examination method.^[9]

With the rapid development of technology in recent years, smart textile products, which are wearable devices, can be preferred for many purpose-specific reasons. Smart textile products are used for purposes such as sensory development of individuals diagnosed with autism spectrum disorder, performance development of visually impaired swimmers, and determination of the risk of diabetic foot ulcers.^[10–12] Smart textile products are developed with the principle of improving and using a determined technical feature. These developed products can detect and react to any external influence or changes in this influence.^[13] Thanks to a series of sensors incorporated into textiles, smart textiles can detect environmental conditions and external stimuli such as light, heat, pressure, electromagnetic waves, sound and ultrasound waves, and movement.^[14]

In this study, “wearable technologies,” “physiotherapy and rehabilitation,” “physical activity,” “sleep,” and “smart textiles” keywords have been searched in Pubmed, Google Scholar, Cochrane, and Scopus databases between 2001 and 2023. Wearable devices used in central nervous system diseases, athletes, sleep, physical activity, and weight control as well as using smart textiles in rehabilitation have been included in our study. This study aims to reveal wearable technologies used in the field of physiotherapy and rehabilitation by making use of current literature and contributing to the greater use of these technologies in new scientific studies.

Clinical and Research Consequences

Wearable Technologies in Central Nervous System Diseases

Central nervous system diseases are a broad category of neurological disorders that cause reduced motor abilities and are considered one of the largest public health problems.

^[15] Central nervous system diseases include Alzheimer's, Parkinson's, brain tumors, stroke, and other important diseases that threaten human life.^[16] Wearable technologies provide significant benefits in the treatment of central nervous system diseases such as Parkinson's, Alzheimer's, and hemiplegia.^[3]

ActiGraph

ActiGraph is a wearable device that can measure step count, activity intensity, total sleep time, sleep efficiency, waking after sleep onset, and total number of movements. Offering USB and Bluetooth® Smart functions, this device captures and records high-resolution human movements thanks to its 3-axis accelerometer. It is quite light (19 g) and provides easy use. When the literature is examined, it is seen that it is used in studies on determining the number of steps and fatigue level in individuals diagnosed with Multiple Sclerosis (MS).^[17,18]

Biovotion everion

Biovotion Everion is a wearable device that can non-invasively collect real-time physiological data such as heart rate, respiratory rate, blood oxygenation, and skin temperature. It is usually worn on the upper arm. In addition, other clinical and non-clinical parameters that can be measured by this device are movement intensity, number of steps, energy consumption, sleep quality, and heart rate variability. It is one of the devices used to determine fatigue in individuals diagnosed with MS.^[19]

Fitbit flex

Fitbit Flex is a device that automatically tracks daily activity count, distance, calories burned, and sleep. The elastomer

material used in this watch is similar to that used in many sports watches, and it is flexible and durable. It also has a 3-axis accelerometer and vibration motor. It also has the feature of losing 7 days of detailed movement data minute by minute. It was used in a study on disability assessment in individuals diagnosed with MS.^[20]

Keeogo exoskeleton

Keeogo exoskeleton has limited walking endurance or mobility. It is a wearable device designed for people who have difficulty climbing stairs or carrying objects. This wearable robotic device also has a strap system that is significantly different from other wearable devices.^[21]

Wolk hip airbag

Wolk hip airbag is a wearable device that aims to prevent hip fractures and other injuries in the event of a fall. It is ergonomically important as users can wear it comfortably under their clothes. The device has six motion sensors and thanks to these sensors, it records every movement 500 times per second. Thanks to its software, this device processes all data and constantly determines whether the user has fallen or not. When its advanced algorithm detects a fall, the belt inflates and prevents the user from falling and getting hurt. Its use can be very beneficial, especially for patients diagnosed with stroke and the elderly population whose fall risk needs to be determined. The subject of studies is to reduce the fear of falling in individuals with stroke during rehabilitation and post-treatment processes.^[22]

Wearable movement sensor

The wearable motion sensor consists of six sensors that detect the general movements of babies aged 12–16 weeks. Sensors weighing 10.25 g are placed so that they do not hinder the baby's natural movements or behavioral state. It is effective in early diagnosis of cerebral palsy and neurodevelopmental disorders and can improve lifelong health outcomes for at-risk infants.^[23]

Shimmer 3

Shimmer 3 is a device with a gyroscope, magnetometer, three-axis accelerometer, and a processor capable of 3D motion estimation. In addition, it demonstrates Bluetooth data streaming functionality. All signals can be measured simultaneously and in real-time. There is a study used to evaluate spasticity in individuals diagnosed with cerebral palsy. It may be an alternative method to evaluation methods of spasticity in the field of physiotherapy and rehabilitation.^[24]

Empatica E4

Empatica E4 is a wearable device that allows the measurement of sympathetic nervous system activity, skin temperature, and heart rate. It has Bluetooth feature, battery, and internal memory that can record up to 60 h. The device weighs 25 g. Thanks to its 3-axis accelerometer, it can detect motion-sensitive activities. There are devices currently used in individuals diagnosed with epilepsy. With the development of technology, it is thought that the use of these devices will increase by both patients and their relatives.^[25,26]

Biometrics-differential sensors/SX230

The biometrics-differential sensors/SX230 device can capture signals during both static and dynamic activities and is simple to use without the need for cream or gel on the skin. It can be easily used with an electromyography (EMG) device in small muscles or large muscle groups. It was used in a study analyzing daily EMG results in women with Parkinson's disease.^[27]

Parkinson's KinetiGraph

Parkinson's KinetiGraph consists of a small device worn on the wrist to collect data over 6–10 days. This smartwatch is used to evaluate symptoms such as tremors, bradykinesia, and dyskinesia during activity in patients with Parkinson's disease.^[28]

Wearable Technologies for Athletes

Xsens MVN Systems

Xsens MVN systems are systems that analyze the lower extremity and pelvis kinematics of football players when they shoot. Unlike videography methods used as the gold standard in the literature for measuring the biomechanics of high-speed movements, it is easy and inexpensive to use in the field. When the literature is examined, it is seen that it is used as a safe device in motion analysis measurements. Since this system is small and easily adaptable to the body, it seems to be advantageous over other methods in evaluating the athlete's field performance.^[29]

ISWIM System

The ISWIM system is designed for swimmers. It has sensors that examine the dynamic movements of athletes and provide the opportunity to instantly correct incorrect kinematics with vibratory feedback. It provides instant feedback to the athlete by measuring swimmers' roll angles (medial-lateral rotation) and stroke speeds during competition or training. While the swimmer receives single vibration feedback when swimming at a normal roll angle,

and double vibration feedback if swimming at an extreme roll angle. In case of insufficient angulation, they do not receive any vibratory feedback.^[30]

OPAL, APDM

OPAL, APDM wearable sensors are a system that measures whole-body kinematics. These sensors, which use gyroscopes and accelerometers, can synchronize up to 24 units wirelessly. This device provides real-time results to its users with a delay of <30 ms. Apart from its use mentioned in the central nervous system diseases section, it is also preferred by athletes. It allows both laboratory and sports-specific in-field use. Although there are many studies in the literature, a study conducted on swimmers stated that, unlike video analysis methods, both in-water and out-of-water analyses can be performed simultaneously and the method is safe.^[31]

Sleep Wearable Devices

Ōura Ring

The Ōura ring was developed to estimate sleep parameters, heart rate variability, respiratory rate, and intensity of physical activity. It is a sleep monitoring device that uses acceleration and gyroscope data, photoplethysmogram signal, and body temperature. The Ōura ring has been shown to have 96% sensitivity in detecting sleep and 65%, 51%, and 61% accuracy in detecting “light sleep,” “deep sleep,” and “rapid eye movement” (REM sleep), respectively.^[32,33]

Vivosmart® 5 Garmin Wristband

Vivosmart® 5 Garmin Wristband is a wearable wristband that collects information about falling asleep, waking from sleep, time spent awake, and duration of stay in basic sleep stages. It identifies sleep durations using a combination of sleep stages, heart rate, heart rate variability, and body movement data. In addition, the bracelet monitors your breathing rate and blood oxygen saturation levels throughout the night. For deep sleep, the nominal sensitivity is 0.56, and the specificity and accuracy values are 0.92 and 0.87, respectively.^[34]

Whoop Bracelet

WHOOP is a wearable device that provides measurements of heart rate, heart rate variability, and sleep staging. This device has an accelerometer to track sleep. It also calculates the user's heart rate variability through the optical sensor. It provides information about the time it takes to fall asleep, time spent asleep, sleep efficiency, number of conscious and/or unconscious awakenings during the night, and

position changes. It has been demonstrated that it is a device that can be used to estimate the stages of sleep in cases where PSG is not performed.^[35,36]

Dreem Headband

The Dreem Headband includes five EEG electrodes, a three-dimensional accelerometer, and a pulse oximeter embedded in the device. The Dreem device has proven capacity to precisely monitor sleep-related physiological signals and accurately process them into sleep stages.^[37]

Owlet Sock

The owlet sock is a miniature, wireless smart sock worn by newborn babies as a sock. It provides information to parents and healthcare professionals about the baby's oxygen saturation, skin temperature, heart rate, and sleep patterns. Owlet sock can detect sleep apnea observed in newborns.^[38]

Go2sleep

Go2sleep is a wearable sleep measurement device. It estimates the depth of sleep by measuring the user's pulse and blood oxygen levels through finger sensors. It has the feature of warning the user in case of sleep apnea.^[39]

Wearable Devices for Physical Activity and Weight Control

Modius

Modius is a wearable device that activates the hypothalamus, increases metabolic rate, and burns fat by stimulating the vestibular nerve. This headphone-shaped device is preferred by adults for weight loss and weight management. For weight loss purposes, it is recommended to use it 5 days a week, in 45-min sessions, for a total of 3 months.^[40]

Lumo Lift

Lumo Lift is a wearable device with a sensor that is usually worn just below the clavicle. It provides information about the user's posture by sending vibration alerts. It also measures steps, distance, and calorie expenditure. The collection and monitoring of data is transferred to an application connected to the smartphone. It is known that posture training significantly reduces lower back pain.^[41,42]

Moticon

Moticon is a wearable device that examines the correct pressure distribution and forces of the movements of patients and athletes with sensor insoles located on the sole of the shoe. It is linked to an application through which data are transferred to the smartphone. In this way, data can be analyzed from a smartphone. There are two types of

sensor insoles. The Moticon OpenGo model has proven to be applicable in both clinical and research settings to evaluate temporal, force, and balance parameters during different types of movement compared to other sensor insoles.^[43]

Smart Textiles Used in Physiotherapy

Adrenaline Dress

Adrenalin Dress can measure a person's body temperature, adrenaline, and stress level. It can detect and analyze the user's sweating, breathing patterns, and body temperature through sensors.^[44,45]

Beagle

Beagle is a hooded scarf designed to stimulate or diminish the senses of its wearer. The device features sound reduction, audio integration, aromatic stimulation, and haptic personalization technologies. This scarf consists of hood pockets. It allows the user to change their sensory experience with the touch of a button. It can be given as an example of smart textile products designed for individuals diagnosed with autism spectrum disorders.^[10]

Nadi X Yoga Pants

Nadi X Yoga Pants have haptic sensors in the fabric layer in the ankle, knee, and hip area. It guides through vibration and sound using sensors embedded in the pants. The application includes a library of yoga poses, visuals, and audio data. It has Bluetooth connectivity and an iOS operating system. The trousers connect to the mobile application through Bluetooth and thus guide the user visually and audibly. In this way, it can instantly warn if something is missing or goes wrong in the position.^[46]

Blind Cap

The Blind Cap is a smart cap developed for visually impaired swimmers. It informs visually impaired swimmers that they have reached the end of the line and need to turn back. It connects through Bluetooth with any Android-based smartphone. The smart bone also makes it easier to make connections between athletes and coaches decently. The trainer ensures that the swimmer receives the necessary warning only when he presses a button on the phone screen. Thus, athletes do not need to touch their heads or bodies to make a turn with a simple mechanism. This smart test product can be preferred to improve the performance of disabled swimmers.^[11]

Hexoskin Pro Kit

Hexoskin Pro Kit is used for advanced physiological health monitoring, longitudinal clinical studies, stress monitoring,

and performance training. It is compatible with iOS and Android operating systems. It connects through Bluetooth and can record for more than 36 h. Hexoskin can determine pulse, heart rate variability, respiratory rate, respiratory volume, and activity level (steps, cadence, and calories). When the literature was examined, it was stated that this device was reliable for determining heart rate, submaximal exercise, and maximum exercise. It has also been found to be valid for measuring tidal volume and minute ventilation during submaximal exercise.^[47]

BalanceBelt

BalanceBelt is designed to provide tactile feedback regarding body position for individuals with balance problems. This device contains several small vibration motors and an accelerometer. The device provides warnings about the user's body position with vibrating feedback. In this way, the user can correct his posture and improve his balance. It is thought that it may serve as an effective aid to improve the quality of life in patients with severe bilateral vestibular loss.^[48]

Conclusion

Wearable technologies are an area where the application area has increased and significant developments have occurred in the last decade. Various wearable solutions are being implemented to assist clinicians in assessing and diagnosing, using commercially available devices or developing custom systems. In addition to contributing to the monitoring, improvement, and sustainability of health, these devices also enable the expansion of the health-related database.

These technologies are commonly used in many diseases' diagnoses such as epilepsy and sleep disorders, in the follow-up of parameters related to physical activity and athlete's health, treatment, and rehabilitation processes of central nervous system diseases such as Alzheimer's and Parkinson's. In addition, smart textile products are used for sensory processing problems of individuals with autism spectrum disorder and to figure out the risk of diabetic foot ulcers. Since wearable technologies have a wide range of uses, health-care professionals recommend these devices to their patients for many different reasons.

Although there are studies conducted with wearable devices in the literature, it seems that they are not included enough in scientific studies. Many randomized controlled scientific studies are needed before the use of these technologies in the diagnosis and treatment processes of diseases can be considered the gold

standard. Since these devices can be used in both diagnosis and treatment processes, studies can be designed for different purposes in scientific studies.

Physicians and physiotherapists play an active role in determining the purpose-specific wearable devices for rehabilitation. Considering wearable technologies' contributions to healthy individuals and patients, these technologies should be used more. In this context, it is very important to increase the awareness level of individuals and provide information about the devices. Smart textile products, like other wearable devices, provide benefits to health-care professionals in the diagnosis and treatment process. It can be predicted that smart textile products will be given more space in scientific studies in the coming years.

This study is important in terms of revealing wearable technologies used for different purposes in the field of physiotherapy and rehabilitation, user-specific wearable device selection, providing necessary information about the devices, and contributing to the greater use of these technologies in new scientific studies.

Disclosures

Peer-review: Externally peer-reviewed.

Conflict of Interest: All authors declared no conflict of interest.

Financial Disclosure: The authors declared that this study received no financial support.

Authorship Contributions: Concept – A.V.Ö., A.K.K., M.K.A.; Design – A.V.Ö., M.K.A., M.O.K.; Supervision – A.V.Ö., O.Ş., A.K.K.; Funding – M.O.K., M.K.A.; Data analysis and/or interpretation – A.K.K., M.O.K., H.D.; Literature search – O.Ş., S.A., H.D.; Writing – O.Ş., S.A., H.D.; Critical review – A.V.Ö., O.Ş., S.A.

References

- Demirci Ş. Giyilebilir teknolojilerin sağlık hizmetlerine ve sağlık hizmet kullanıcılarına etkileri. *Anemon Muş Alparslan Üniv Sosyal Bilimler Derg* 2018;6(6):985–92.
- Surantha N, Atmaja P, Wicaksono M. A review of wearable internet-of-things device for healthcare. *Procedia Comput Sci* 2021;179:936–43.
- Lu L, Zhang J, Xie Y, Gao F, Xu S, Wu X, et al. Wearable health devices in health care: Narrative systematic review. *JMIR Mhealth Uhealth* 2020;8(11):e18907.
- Erkiliç CE, Yalçın A. Evaluation of the wearable technology market within the scope of digital health technologies. *Gazi İktisat İşletme Derg* 2020;6(3):310–23.
- Perez AJ, Zeadally S. Recent advances in wearable sensing technologies. *Sensors (Basel)* 2021;21(20):6828.
- Strain T, Wijndaele K, Dempsey PC, Sharp SJ, Pearce M, Jeon J, et al. Wearable-device-measured physical activity and future health risk. *Nat Med* 2020;26(9):1385–91.
- Buyse DJ. Sleep health: Can we define it? Does it matter? *Sleep* 2014;37(1):9–17.
- Hublin C, Partinen M, Koskenvuo M, Kaprio J. Sleep and mortality: A population-based 22-year follow-up study. *Sleep* 2007;30(10):1245–53.
- Chung KF, Yeung WF, Ho FY, Yung KP, Yu YM, Kwok CW. Cross-cultural and comparative epidemiology of insomnia: The Diagnostic and statistical manual (DSM), International classification of diseases (ICD) and International classification of sleep disorders (ICSD). *Sleep Med* 2015;16(4):477–82.
- Koo SH, Gaul K, Rivera S, Pan T, Fong D. Wearable technology design for autism spectrum disorders. *Arch Des Res* 2018;31(1):37–55.
- Oommen J, Bews D, Hassani MS, Ono Y, Green JR. A Wearable Electronic Swim Coach for Blind Athletes. In: 2018 IEEE Life Sciences Conference (LSC). Montreal, QC, Canada: IEEE; 2018. p. 219–22.
- Reyzelman AM, Koelewyn K, Murphy M, Shen X, Yu E, Pillai R, et al. Continuous temperature-monitoring socks for home use in patients with diabetes: Observational study. *J Med Internet Res* 2018;20(12):e12460.
- Tao X. Smart technology for textiles and clothing-introduction and review. In: *Smart Fibres, Fabrics and Clothing*. Cambridge: Woodhead Publishing; 2001. p. 1–6.
- Alemohammad H, editor. *Opto-mechanical Fiber Optic Sensors: Research, Technology, and Applications in Mechanical Sensing*. United Kingdom: Butterworth-Heinemann; 2018. p. 327.
- Locatelli P, editors. *CNS Diseases Monitoring Using Wearable Devices. Monitoring Techniques Based on Inertial Platforms for Patients Affected by Central Nervous System Diseases*. Italy: Università degli Studi di Bergamo; 2021.
- Wei W, Qiu Z. Diagnostics and theranostics of central nervous system diseases based on aggregation-induced emission luminogens. *Biosens Bioelectron* 2022;217:114670.
- Learmonth YC, Motl RW, Sandroff BM, Pula JH, Cadavid D. Validation of patient determined disease steps (PDDS) scale scores in persons with multiple sclerosis. *BMC Neurol* 2013;13:37.
- Ibrahim AA, Flachenecker F, Gaßner H, Rothhammer V, Klucken J, Eskofier BM, et al. Short inertial sensor-based gait tests reflect perceived state fatigue in multiple sclerosis. *Mult Scler Relat Disord* 2022;58:103519.
- Barrios L, Oldrati P, Santini S, Lutterotti A. Recognizing Digital Biomarkers for Fatigue Assessment in Patients with Multiple Sclerosis. In: *Proceedings of the 12th EAI International Conference on Pervasive Computing Technologies for Healthcare; PervasiveHealth*. New York: EAI; 2018.

20. Block VJ, Lizée A, Crabtree-Hartman E, Bevan CJ, Graves JS, Bove R, et al. Continuous daily assessment of multiple sclerosis disability using remote step count monitoring. *J Neurol* 2017;264(2):316–26.
21. McGibbon CA, Sexton A, Jayaraman A, Deems-Dluhy S, Gryfe P, Novak A, et al. Evaluation of the Keeogo exoskeleton for assisting ambulatory activities in people with multiple sclerosis: An open-label, randomized, cross-over trial. *J Neuroeng Rehabil* 2018;15(1):117.
22. Botonis OK, Harari Y, Embry KR, Mummidisetty CK, Riopelle D, Giffhorn M, et al. Wearable airbag technology and machine learned models to mitigate falls after stroke. *J Neuroeng Rehabil* 2022;19(1):60.
23. Redd CB, Barber LA, Boyd RN, Varnfield M, Karunanithi MK. Development of a wearable sensor network for quantification of infant general movements for the diagnosis of cerebral palsy. *Annu Int Conf IEEE Eng Med Biol Soc* 2019;2019:7134–9.
24. Choi S, Shin YB, Kim SY, Kim J. A novel sensor-based assessment of lower limb spasticity in children with cerebral palsy. *J Neuroeng Rehabil* 2018;15(1):45.
25. Nasser M, Nurse E, Glasstetter M, Böttcher S, Gregg NM, Nandakumar AL, et al. Signal quality and patient experience with wearable devices for epilepsy management. *Epilepsia* 2020;61 Suppl 1:S25–35.
26. Mancini M, Horak FB. Potential of APDM mobility lab for the monitoring of the progression of Parkinson's disease. *Expert Rev Med Devices* 2016;13(5):455–62.
27. Roland KP, Jones GR, Jakobi JM. Daily electromyography in females with Parkinson's disease: A potential indicator of frailty. *Arch Gerontol Geriatr* 2014;58(1):80–7.
28. Guan I, Trabilsy M, Barkan S, Malhotra A, Hou Y, Wang F, et al. Comparison of the Parkinson's KinetiGraph to off/on levodopa response testing: Single center experience. *Clin Neurol Neurosurg* 2021;209:106890.
29. Blair S, Duthie G, Robertson S, Hopkins W, Ball K. Concurrent validation of an inertial measurement system to quantify kicking biomechanics in four football codes. *J Biomech* 2018;73:24–32.
30. Li R, Cai Z, Lee W, Lai DT. A wearable biofeedback control system based body area network for freestyle swimming. *Annu Int Conf IEEE Eng Med Biol Soc* 2016;2016:1866–9.
31. Fantozzi S, Giovanardi A, Magalhães FA, Di Michele R, Cortesi M, Gatta G. Assessment of three-dimensional joint kinematics of the upper limb during simulated swimming using wearable inertial-magnetic measurement units. *J Sports Sci* 2016;34(11):1073–80.
32. de Zambotti M, Rosas L, Colrain IM, Baker FC. The sleep of the ring: Comparison of the ŌURA sleep tracker against polysomnography. *Behav Sleep Med* 2019;17(2):124–36.
33. Stevens S, Siengsukon C. Commercially-available wearable provides valid estimate of sleep stages 2019;15(suppl):p3. 6-042.
34. Grandner MA, Lujan MR, Ghani SB. Sleep-tracking technology in scientific research: Looking to the future. *Sleep* 2021;44(5):zsab071.
35. Miller DJ, Lastella M, Scanlan AT, Bellenger C, Halson SL, Roach GD, et al. A validation study of the WHOOP strap against polysomnography to assess sleep. *J Sports Sci* 2020;38(22):2631–6.
36. Miller DJ, Sargent C, Roach GD. A validation of six wearable devices for estimating sleep, heart rate and heart rate variability in healthy adults. *Sensors (Basel)* 2022;22(16):6317.
37. Arnal PJ, Thorey V, Debellemanniere E, Ballard ME, Bou Hernandez A, Guillot A, et al. The dreem headband compared to polysomnography for electroencephalographic signal acquisition and sleep staging. *Sleep* 2020;43(11):zsaa097.
38. Dangerfield MI, Ward K, Davidson L, Adamian M. Initial experience and usage patterns with the owlet smart sock monitor in 47,495 newborns. *Glob Pediatr Health* 2017;4:2333794X17742751.
39. Cay G, Ravichandran V, Sadhu S, Zisk AH, Salisbury A, Solanki D, et al. Recent Advancement in Sleep Technologies: A Literature Review on Clinical Standards, Sensors, Apps, and AI methods. New Jersey: IEEE Access; 2022.
40. Coates McCall I, Lau C, Minielly N, Illes J. Owning ethical innovation: Claims about commercial wearable brain technologies. *Neuron* 2019;102(4):728–31.
41. Park S, Hetzler T, Hammons D, Ward G. Effects of biofeedback postural training on pre-existing low back pain in static-posture workers. *J Back Musculoskelet Rehabil* 2018;31(5):849–57.
42. Ruiz-Malagón EJ, Delgado-García G, Castro-Infantes S, Ritacco-Real M, Soto-Hermoso VM. Validity and reliability of NOTCH® inertial sensors for measuring elbow joint angle during tennis forehand at different sampling frequencies. *Measurement* 2022;201:111666.
43. Stöggel T, Martinier A. Validation of Moticon's OpenGo sensor insoles during gait, jumps, balance and cross-country skiing specific imitation movements. *J Sports Sci* 2017;35(2):196–206.
44. Porta JP. Validating the Adidas miCoach and Nike+ Sport Kit for Estimating Pace, Distance, and Energy Expenditure During Over-ground Exercise. Texas: The University of Texas at El Paso; 2013.
45. Değerli NG. Moda endüstrisinin giyilebilir teknoloji tasarımları. *Uluslararası Bilimsel Araştırmalar Derg* 2019;4(1):50–65.
46. Long C, Jo E, Nam Y. Development of a yoga posture coaching system using an interactive display based on transfer learning. *J Supercomput* 2022;78(4):5269–84.

-
47. Smith CM, Chillrud SN, Jack DW, Kinney P, Yang Q, Layton AM. Laboratory validation of hexoskin biometric shirt at rest, submaximal exercise, and maximal exercise while riding a stationary bicycle. *J Occup Environ Med* 2019;61(4):e104–11.
 48. Kingma H, Felipe L, Gerards MC, Gerits P, Guinand N, Perez-Fornos A, et al. Vibrotactile feedback improves balance and mobility in patients with severe bilateral vestibular loss. *J Neurol* 2019;266(Suppl 1):19–26.