

# Green MIP-Based Electrochemical Sensing Platform for Environmental Ivermectin Analysis

Zeynep Aydemir, Beril S. Kaya, Setareh Dorreh, Abdullah Al Faysal, Taner Erdoğan, and Ayşegül Gölçü\*

Cite This: *ACS Omega* 2026, 11, 27050–27062

Read Online

ACCESS |



Metrics &amp; More




Article Recommendations



Supporting Information

**ABSTRACT:** Ivermectin (IVM), a macrocyclic lactone derived from *Streptomyces avermitilis*, is widely recognized as a “wonder drug” for its broad-spectrum efficacy against internal and external parasites in human and veterinary medicine. Owing to its potent pharmacological activity, precise quantification of IVM is essential for therapeutic monitoring and dose optimization. In this study, we report the design of a novel electrochemical sensor based on molecularly imprinted polymer (MIP) technology, specifically tailored for the selective detection of IVM. The sensor was fabricated via an electropolymerization strategy employing methacrylic acid (MAA) as the functional monomer and aniline as the comonomer in phosphate-buffered saline (PBS, pH 7.0). To the best of our knowledge, this represents the first electropolymerization-based MIP sensor developed for IVM determination. The resulting MAA-IVM@MIP/GCE sensor was thoroughly characterized using cyclic voltammetry (CV), electrochemical impedance spectroscopy (EIS), Fourier-transform infrared spectroscopy (FTIR), and scanning electron microscopy (SEM). Electrochemical detection was achieved through an indirect redox-probe approach with 5.0 mM  $[\text{Fe}(\text{CN})_6]^{3-/4-}$ , providing a wide linear range ( $1 \times 10^{-12}$ – $1 \times 10^{-11}$  M) and remarkably low limits of detection (LOD:  $2.91 \times 10^{-13}$  M) and quantification (LOQ:  $9.71 \times 10^{-13}$  M). The sensor demonstrated high sensitivity, reproducibility, and selectivity, clearly distinguishing IVM from structurally related compounds. It maintained strong analytical performance in pharmaceutical formulations, biological matrices, and environmental samples such as tap water and soil, showing minimal matrix interference. These results confirm the platform’s robustness and applicability. Density functional theory (DFT) calculations were performed to evaluate template–monomer interactions and determine the optimal template:monomer ratio for the MIP-based sensor. The results revealed that the 1:1 complex exhibited the most favorable binding characteristics, consistent with the experimental findings. In addition, the sensor fabrication strategy was designed in accordance with green analytical chemistry principles. The electropolymerization process was performed in aqueous phosphate-buffered saline under mild conditions without the use of excessive cross-linkers or hazardous reagents. The approach minimizes organic solvent consumption, reduces energy requirements, and enables sensor reusability, thereby contributing to a sustainable and environmentally responsible analytical platform. Overall, this cost-effective, scalable, and environmentally conscious electrochemical sensor provides a practical tool for reliable IVM monitoring and has strong potential for clinical diagnostics, pharmacokinetics, and pharmaceutical quality control.



## 1. INTRODUCTION

Originally isolated from the actinomycete *Streptomyces avermitilis*, ivermectin (IVM) belongs to the macrocyclic lactone family of compounds.<sup>1</sup> It marked a breakthrough in veterinary medicine as the first drug of its kind to be used against parasitic infections, and has since become the most extensively used agent for treating both internal and external parasites in animals.<sup>2</sup> Beyond its antiparasitic properties, IVM has also been described as a “wonder drug” due to its broad range of biological activities, including antimicrobial, antiviral, and anticancer effects.<sup>3</sup> Notably, its antiviral action extends to both DNA and RNA viruses. IVM inhibits nuclear transport by targeting the importin  $\alpha/\beta$ 1 heterodimer, which is essential for the translocation of viral proteins into the host cell nucleus, thereby disrupting viral replication processes.<sup>4</sup> Given its broad biological profile, IVM was even investigated during the

COVID-19 pandemic as a potential treatment option. Although later studies yielded inconclusive results, their widespread use during that period highlights the interest in repurposing established drugs during global health emergencies.<sup>5</sup>

As IVM exhibits multiple biological effects, ensuring accurate methods for its detection and quantification has become a key focus in both scientific and medical fields. Over

Received: January 18, 2026

Revised: March 9, 2026

Accepted: April 9, 2026

Published: April 30, 2026



the years, various detection techniques have been developed and refined to meet this need. Early detection approaches employed thin-layer chromatography (TLC), enabling qualitative identification of IVM as a fluorescent derivative in cattle dung, with a limit of detection (LOD) as low as  $\leq 40$  ng/g.<sup>6</sup> Subsequent advancements led to competitive enzyme immunoassays developed for bovine liver samples, offering a LOD as low as 1.6 ng/g and demonstrating strong correlation with high-performance liquid chromatography (HPLC) results.<sup>7</sup> For food safety monitoring, immunoaffinity column cleanup combined with reversed-phase HPLC allowed IVM quantification in swine liver with a LOD of 2  $\mu$ g/kg and recovery rates ranging from 85% to 102%.<sup>8</sup> In aquaculture, an HPLC method with fluorescence detection enabled precise measurement of IVM residues in Atlantic salmon tissues, with a sensitivity of 1 ng/g, liver being the tissue with the highest accumulation.<sup>9</sup> Recent efforts have focused on immunochemical methods: a monoclonal antibody-based indirect competitive enzyme-linked immunosorbent assay (ic-ELISA) achieved a LOD of 0.77 ng/mL, while a lateral-flow immunochromatographic assay (ICA) provided a visual detection limit of 25 ng/mL and a scanner-based LOD of 2.9 ng/mL.<sup>10</sup> These methods provide accurate and sensitive results, but their application is restricted due to the high cost, slow processing, and requirement for expert handling and complex tools.

Compared to the aforementioned techniques, electrochemical methods offer several advantages, including lower limits of detection, rapid analysis time, cost-effectiveness, and suitability for miniaturization.<sup>11</sup> These features make them particularly attractive for sensitive, on-site, and real-time monitoring of IVM residues in various matrices. Furthermore, employing surface modification techniques can significantly improve the electrochemical performance of working electrodes, enhancing both sensitivity and selectivity. One promising approach involves the use of molecularly imprinted polymers (MIPs), which are cost-effective, durable, and demonstrate excellent performance and stability under a range of conditions.<sup>12</sup> Thanks to their mechanical strength, thermal stability, and ease of transport, MIPs have gained significant interest as analytical tools in pharmaceutical applications. These polymers can be synthesized with minimal expense from readily available materials, offering notable sensitivity that supports their integration into a variety of applications—most notably, electrochemical sensor technologies.<sup>13</sup> To fabricate an MIP on the surface of an electrode, functional monomers are combined with a target template and a cross-linking agent to form a structured polymer matrix. Once polymerization is complete, the template is removed, leaving behind specific binding sites that are complementary in shape, size, and chemical functionality to the original molecule.<sup>14,15</sup> The rigidity of the polymer matrix is also recognized as a key parameter governing MIP performance, as higher structural stiffness generally improves the stability and fidelity of the imprinted cavities, enhancing selectivity, while overly rigid matrices may limit analyte diffusion and binding kinetics. This balance between mechanical stability and accessibility has been highlighted as an important design consideration in molecular imprinting.<sup>16</sup> In recent years, computational approaches have become valuable tools in the design of MIPs, particularly for predicting interactions between monomers and templates in the prepolymerization phase.<sup>17</sup> Among these methods, density functional theory (DFT) has proven especially useful in identifying optimal functional monomers based on calculated

interaction energies with the target template. DFT has been effectively employed in selecting both monomers and solvents for MIP synthesis using templates.<sup>18</sup>

In recent years, the development of analytical methods has increasingly been guided by the principles of green analytical chemistry (GAC), which aim to minimize environmental impact while maintaining high analytical performance.<sup>19</sup> Key green profile parameters include the reduction of hazardous reagents, minimization of organic solvent consumption, lower energy requirements, and the design of reusable and waste-reducing analytical platforms. Electrochemical sensing technologies, particularly those based on MIPs, align well with these principles due to their low reagent consumption, mild operating conditions, and potential for miniaturization and reuse. Therefore, integrating green chemistry considerations into sensor design not only enhances environmental sustainability but also improves the practicality and scalability of analytical methods, and such aspects can be systematically evaluated using modern greenness assessment metrics.<sup>20</sup>

Despite the significant progress achieved with chromatographic, immunochemical, and spectroscopic techniques for IVM determination, many of these approaches still rely on expensive instrumentation, labor-intensive sample preparation, and centralized laboratory settings, which limit their suitability for rapid and on-site monitoring. Electrochemical sensing platforms, particularly those incorporating MIPs, have emerged as technologically relevant alternatives due to their high sensitivity, portability, and cost-effectiveness. In this context, integrating computationally guided MIP design with electrochemical transduction represents a promising strategy to enhance selectivity while maintaining operational simplicity. Therefore, the development of new MIP-based electrochemical sensors tailored for IVM detection remains highly relevant within the current landscape of analytical technologies.

In this study, we developed and optimized a novel electrochemical sensing platform by combining MIP technology with electrochemical techniques for the selective and sensitive detection of IVM. The polymer matrix was prepared using methacrylic acid (MAA) as the functional monomer and aniline as the comonomer in a phosphate-buffered saline (PBS, pH 7.0) solution. The electro polymerization was carried out directly on the surface of a glassy carbon electrode (GCE) using cyclic voltammetry, yielding a stable, imprinted polymeric film.

The morphological and electrochemical properties of the synthesized MIP were characterized using scanning electron microscopy (SEM), Fourier-transform infrared spectroscopy (FTIR), electrochemical impedance spectroscopy (EIS), and cyclic voltammetry (CV). The resulting MAA-IVM@MIP/GCE exhibited strong molecular recognition capability toward IVM, even in the presence of structurally similar compounds. Compared to previously reported sensors, the developed platform showed enhanced sensitivity and selectivity. Moreover, its successful application to pharmaceutical and environmental samples highlights its practical applicability. Overall, this electrochemical sensor provides a reliable, efficient tool for accurate detection of IVM, contributing considerably to drug monitoring and quality control efforts in the pharmaceutical and veterinary fields.

## 2. EXPERIMENTAL SECTION

### 2.1. Reagents and Chemicals

In this study, IVM was obtained from Istanbul University, Drug Research Center, and the tablet form was obtained upon request from Deva Holding. Additional chemicals were sourced from Sigma-Aldrich (Darmstadt, Germany), including methacrylic acid (MAA,  $\geq 99.0\%$ ), aniline ( $\geq 99.5\%$ ), sodium phosphate monobasic ( $\geq 99.0\%$ ), sodium phosphate dibasic ( $\geq 99.0\%$ ), dopamine (DOP,  $99.0\%$ ), sodium sulfate ( $\text{Na}_2\text{SO}_4$ ,  $>97.0\%$ ), uric acid (UA,  $\geq 99.0\%$ ), potassium nitrate ( $\text{KNO}_3$ ,  $\geq 99.0\%$ ), acetonitrile (ACN,  $99.9\%$ ), ascorbic acid (AA,  $\geq 99.0\%$ ), potassium ferrocyanide and ferricyanide ( $\geq 99.0\%$ ), paracetamol (PAR,  $\geq 99.0\%$ ), acetone ( $99.5\%$ ), magnesium chloride ( $\text{MgCl}_2$ ,  $\geq 98.0\%$ ), potassium chloride (KCl,  $\geq 99.0\%$ ), 2-hydroxy-2-methylpropionophenone ( $\geq 97.0\%$ ), acetic acid (HAc,  $99.0\%$ ), methanol (MeOH,  $99.9\%$ ), ethanol (EtOH,  $99.5\%$ ), hydrochloric acid ( $37\%$ ), and sodium hydroxide ( $>97\%$ ). All substances were employed without any further purification or processing.

Stock solutions were prepared as follows: IVM was dissolved in ethanol to obtain a 10 mM solution; MAA was dissolved in deionized water to a concentration of 10 mM; and an equimolar 5 mM solution of potassium ferrocyanide/ferricyanide was prepared in 0.1 mM KCl. Each solution underwent 10 min of sonication using an ultrasonic bath and was then stored at  $4^\circ\text{C}$  until use. All preparations were conducted using ultrapure water maintained at  $25^\circ\text{C}$ .

### 2.2. Equipment/Apparatus

Electrochemical measurements, including CV and differential pulse voltammetry (DPV), were conducted using a potentiostat/galvanostat (AUTOLAB PGSTAT302N, Metrohm, Utrecht, Netherlands) controlled through NOVA software version 2.1.6. EIS measurements were conducted using a PalmSens potentiostat (PalmSens BV, Houten, Netherlands) with PSTrace software version 5.12.1031. The electrochemical setup employed a conventional three-electrode system, consisting of an Ag/AgCl reference electrode (3 M KCl), a platinum wire serving as the counter electrode, and a glassy carbon electrode (GCE, 3.0 mm diameter) modified with either a MIP or a nonimprinted polymer (NIP) as the working electrode.

Reagents were accurately measured with an analytical balance (Ohaus Corporation, Shanghai, China). Drug extraction and rebinding processes were performed using a thermoshaker (Biosan TS-100, Riga, Latvia). Additional instruments used throughout the experimental procedures included an ultrasonic bath (JP Selecta, Barcelona, Spain) and a vortex mixer (ISOLAB Laborgeräte GmbH, Germany).

For morphological analysis of the polymer films, SEM was performed using a Tescan GAIA3 SEM-FIB system (FEI Quanta FEG 250, USA). To examine the chemical structure of the polymeric materials, ATR-FTIR spectroscopy was performed using a Shimadzu IRSpirit-T spectrometer (Shimadzu, Japan), spanning the mid-infrared range of  $4000\text{--}500\text{ cm}^{-1}$ .

### 2.3. Preparation Processes of EP-IVM@MIP/GCE and EP-IVM@NIP/GCE Sensors

Before electropolymerization, the GCE was cleaned by sonicating for 15 min in a 1:1 (v/v) solution of double-distilled water and methanol. The electrode surface was then polished with an alumina slurry, rinsed thoroughly with double-distilled water, and air-dried at room temperature.

To prepare the electropolymerization (EP) solution, a mixture containing 100  $\mu\text{L}$  of IVM (10.0 mM), 100  $\mu\text{L}$  of methacrylic acid (MAA, 10.0 mM), 500  $\mu\text{L}$  of aniline (0.1 M), and 4300  $\mu\text{L}$  of phosphate-buffered saline (PBS, pH 7.0) was used. A total of 5 mL of this solution was transferred into an electrochemical cell, where the GCE was immersed to carry out the EP process. Electropolymerization was performed by CV over 10 cycles within the potential window of  $-0.2$  to  $+1.2\text{ V}$  at a scan rate of  $50\text{ mV/s}$ . After polymer formation, the electrode was rinsed thoroughly with distilled water.

To remove the template molecule (IVM), the modified electrode was immersed in an EtOH:ACN mixture (2:1, v/v) and incubated on

a thermoshaker at  $25^\circ\text{C}$  and 650 rpm for 10 min. Subsequently, rebinding of IVM ( $5 \times 10^{-11}\text{ M}$ ) to the vacant imprinted sites was achieved by incubating the electrode under the same thermoshaking conditions for 15 min.

Control experiments were conducted using NIP, which were synthesized under identical conditions but without the addition of IVM. The electrochemical performance of the sensors was evaluated using a 5.0 mM  $[\text{Fe}(\text{CN})_6]^{3-/4-}$  redox probe by CV and DPV, and the results were compared with those obtained for the nonmolecular imprinted polymer (NIP)-modified GCE.

### 2.4. Analysis of IVM in Tablet Form and Serum Sample Applications

The practical performance of the fabricated sensor was evaluated using a commercial IVM tablet formulation. Five tablets were accurately weighed and crushed into a fine powder using a mortar. Based on the labeled content (3 mg IVM per tablet), an appropriate amount of the powdered sample was calculated and dissolved in ethanol to prepare a 1.0 mM IVM stock solution. This solution was then sonicated for 15 min to ensure complete dissolution. Serial dilutions were subsequently prepared from the stock solution to construct a calibration curve and conduct recovery studies, validating the sensor's performance in detecting IVM within the pharmaceutical matrix.

For serum analysis, 0.5 mL of a 1 mM IVM solution was mixed with 2.7 mL of acetonitrile and 1.8 mL of drug-free human serum (previously stored at  $-20^\circ\text{C}$ ) in a centrifuge tube. To remove proteins, the sample was centrifuged at 5000 rpm for 20 min. The resulting clear supernatant was carefully collected and diluted with ethanol for further calibration and recovery studies. All measurements were carried out in triplicate to assess the reproducibility and precision of the analytical method, and the relative standard deviation (RSD) was calculated for each data set.


### 2.5. Recovery Studies of IVM from Environmental Matrices

The aqueous stock solution was prepared by diluting 1 mL of a 1 mM drug solution with 9 mL of tap water to obtain a 0.1 mM mixture, which was then used for recovery tests at three concentration levels. Tap water samples were spiked with IVM at  $1 \times 10^{-12}\text{ M}$ ,  $2.5 \times 10^{-12}\text{ M}$ , and  $5 \times 10^{-12}\text{ M}$ , after which standard IVM solutions of  $6.5 \times 10^{-12}\text{ M}$ ,  $5 \times 10^{-12}\text{ M}$ , and  $2.5 \times 10^{-12}\text{ M}$  were added to the corresponding samples. The resulting data were used to calculate average recovery (%) and bias (%) to evaluate method accuracy.

Soil samples collected from the Istanbul Technical University campus were dried to homogenize and then heated to  $105^\circ\text{C}$  to remove residual moisture. One gram of the dried soil was combined with 10 mL of deionized water, treated in an ultrasonic bath to enhance analyte release, and then centrifuged to separate the extract. One milliliter of the clear supernatant was diluted to 10 mL to produce the soil extract stock solution. Recovery assessments were performed using this extract by adding standard solutions ( $6.5 \times 10^{-12}\text{ M}$ ,  $5.0 \times 10^{-12}\text{ M}$ , and  $2.5 \times 10^{-12}\text{ M}$ ) to create three fortified levels. The sensor was used to analyze these samples, and the measured concentrations were compared to the expected values to determine recovery and bias. Recoveries between 95.0% and 105.0% were deemed satisfactory, confirming the method's suitability for soil analysis.

### 2.6. Computational Studies

Density functional theory (DFT) calculations were performed to determine the optimum template:monomer ratio for the MIP-based sensor and to investigate possible template-monomer interactions. Initially, the three-dimensional structure of ivermectin was constructed, and a conformational analysis was carried out to identify the most stable conformer. Geometry optimization of this conformer was then performed using B3LYP (Becke, three-parameter, Lee–Yang–Parr hybrid functional)<sup>21,22</sup> functional in combination with the 6–311G(d,p) basis set. Based on the optimized geometry of ivermectin, template-monomer complexes were generated at template:monomer ratios ranging from 1:1 to 1:5. The initial geometries of these complexes were obtained using AutoDock Tools<sup>23</sup> and AutoDock



**Figure 1.** Surface characterization of the modified electrode. SEM micrographs showing the surface morphology of the MIP-modified electrode (A–C) and the NIP-modified electrode (D); ATR–FTIR spectra corresponding to the MIP (E) and NIP (F) films.

Vina.<sup>24</sup> Subsequently, DFT calculations employing the same level of theory were performed to optimize the geometries of each complex and to calculate their corresponding energies ( $\Delta E_{\text{complex}}$ ). In the next step, for each complex, the optimized geometric structure was first modified by removing the monomer unit(s) from the complex, yielding the template-only structure. Energy calculation was then performed at the same level of theory, and the corresponding energy values ( $\Delta E_{\text{template}}$ ) were obtained. The same procedure was subsequently applied by removing the template molecule from the geometry-optimized complex, resulting in a structure composed solely of the monomer units. Again, energy calculation was carried out, and the corresponding energy values ( $\Delta E_{\text{monomer(s)}}$ ) were determined. Finally, the binding energy for each complex was calculated by the following formula:


$$\Delta E_{\text{binding}} = \Delta E_{\text{complex}} - (\Delta E_{\text{template}} + \Delta E_{\text{monomer(s)}}) \quad (1)$$

DFT calculations were performed with the use of Gaussian<sup>25</sup> and GaussView<sup>26</sup> software packages. VeraChem Vconf<sup>27</sup> was used in conformational analysis and Discovery Studio Visualizer<sup>28</sup> was used to visualize the results.

## 3. RESULTS AND DISCUSSION

### 3.1. Surface Characterization

A detailed morphological analysis of the sensor surfaces was conducted using SEM, with the primary objective of investigating surface structural features and differentiating between the MIPs and NIPs. As depicted in Figures 1A–D, the SEM images provided clear evidence of a significant contrast between the MIP and NIP surfaces of the sensor. The MIP-modified surface exhibited a porous, rough morphology,



**Figure 2.** CV (A), DPV (B), and EIS (C) measurements of the GCE recorded at different stages: before polymerization, following polymer formation, after template removal, and after IVM rebinding, using a 5.0 mM  $[\text{Fe}(\text{CN})_6]^{3-/4-}$  solution (for CV and DPV potential scan range:  $-0.2$  to  $+0.8$  V, scan rate:  $0.05$  V/s, and step potential:  $0.01$  V; for EIS, minimum frequency:  $0.1$  Hz, maximum frequency:  $100,000$  Hz, and  $E_{ac}$ :  $0.01$  V).


consistent with the expected imprinting results (Figures 1A–C). These specific structural features indicate the successful formation of recognition cavities designed for the target analyte. The presence of porosity and surface irregularities is critical for enhancing the analyte's binding capacity. Conversely, the NIP surface exhibited a notably smooth and uniform texture, lacking the porous characteristics of the MIPs, a result anticipated by the deliberate omission of template molecules during polymerization (Figure 1D). Ultimately, the SEM analysis confirmed the success of the imprinting process by visually demonstrating distinct morphological differences between the MIP and NIP surfaces.

The functional groups of the chemical components employed in sensor fabrication were characterized by FTIR spectroscopy. Figures 1E and 1F present the FTIR spectra of both MIP (red line) and NIP (blue line) films. A broad absorption band observed in the range of  $3000$ – $3500$   $\text{cm}^{-1}$  is attributed to the stretching vibrations of N–H groups from aniline and the –OH groups from methacrylic acid. The C–H stretching vibrations of CH and  $\text{CH}_2$  groups appear between  $2900$  and  $3000$   $\text{cm}^{-1}$ . A distinct band around  $1700$   $\text{cm}^{-1}$  is indicative of the C=O stretching vibration associated with

methacrylic acid. Additionally, peaks related to the aromatic ring of aniline are observed in the fingerprint region (below  $1600$   $\text{cm}^{-1}$ ), specifically around  $1600$   $\text{cm}^{-1}$  and  $1500$   $\text{cm}^{-1}$  due to C=C stretching vibrations of the aromatic ring, and C–N stretching vibrations around  $1300$   $\text{cm}^{-1}$ . The overall similarity in the FTIR spectra of MIP and NIP films suggests that the basic polymeric backbone formed from methacrylic acid and aniline is consistent between both formulations, with any differences due to the template (IVM) being subtle and not prominently visible as distinct new peaks, but rather potentially as slight shifts or intensity changes if significant interactions occur. These spectral features collectively verify the successful incorporation of functional groups into the polymeric matrix.

### 3.2. Electrochemical Characterization

The electrochemical response of the MAA-IVM@MIP/GCE sensor was assessed by CV at various fabrication stages—namely after polymerization, template extraction, and subsequent rebinding—using a  $5$  mM  $[\text{Fe}(\text{CN})_6]^{3-/4-}$  redox couple in  $0.1$  M KCl. As depicted in Figure 2A, notable variations in anodic and cathodic peak currents were observed throughout the modification process. The bare electrode



**Figure 3.** Calibration plots for IVM recorded with MIP- and NIP-based sensors in (A) standard solutions and (C) spiked serum samples, together with the corresponding DPV voltammograms illustrating sensor responses at different IVM concentrations in (B) standard solutions and (D) spiked serum. All measurements were carried out in 5 mM  $[\text{Fe}(\text{CN})_6]^{3-/4-}$  using the DPV method.

displayed the highest peak currents, owing to the unobstructed electron transfer at its surface. Following electropolymerization, a substantial reduction in peak currents was observed, indicating that the polymer layer impeded electron transfer and confirming successful imprinting. Removal of the IVM template led to the formation of recognition cavities within the polymer matrix, thereby increasing redox activity, as evidenced by higher peak currents than those of the polymer-coated sensor. Finally, upon rebinding of IVM at a defined concentration, a subsequent decrease in peak currents was detected, attributed to the occupation of the imprinted sites by IVM molecules, thus again limiting electron transfer at the MIP-modified electrode surface.

Figure 2B depicts the DP voltammogram responses acquired from a standard analyte solution at different phases of the fabrication process of the MAA-IVM@MIP/GCE sensor. The peak current shows a reduction following the polymer formation, which is then succeeded by an increase after the template is removed, a phenomenon attributed to the creation of imprinted cavities. Following this, there is yet another decline in peak current during the analyte rebinding process, thus confirming the successful fabrication of the sensor and its specific recognition abilities.

The stepwise modification of the electrode was characterized using EIS to evaluate the interfacial charge-transfer resistance ( $R_{ct}$ ). As shown in the Nyquist plots (Figure 2C), the bare electrode displayed an initial  $R_{ct}$  of 155.6  $\Omega$ . Following electropolymerization, the deposition of the polymer coating caused a sharp increase in impedance to 4315  $\Omega$ . After the template removal step, the  $R_{ct}$  value decreased significantly to 656.5  $\Omega$ , reflecting improved conductivity through the newly created molecular footprints. Finally, the rebinding of IVM led to a renewed increase in impedance ( $R_{ct} = 1332 \Omega$ ), as the analyte filled the imprinted pores and restricted electron transfer.

### 3.3. Optimization Parameters

**3.3.1. Monomer-to-Template Ratio.** The monomer-to-template ratio is a key factor in producing stable, efficient polymers, as it directly governs the bonding interactions between the functional monomer and the template. Choosing a suitable monomer helps reduce nonspecific interactions while promoting selective binding with the target analyte. In this work, MAA was employed as the functional monomer. To determine the optimal molar ratio, changes in peak current ( $\Delta I_1$ ) before and after template extraction were examined at

ratios ranging from 1:1 to 1:5, as shown in Figure S1A. The 1:1 ratio yielded the most favorable outcome, resulting in a robust polymer framework with distinct, selective cavities. At higher ratios, this led to heterogeneous monomer distribution, disrupting the formation of selective binding sites. Thus, the 1:1 ratio was identified as the most effective balance, ensuring efficient template removal and polymerization, as evidenced by the highest  $\Delta I_1$  value.

**3.3.2. Number of Cycles for EP.** The durability and effectiveness of the polymeric film are strongly influenced by the number of scans applied during the EP process. After selecting suitable monomers and their ratios, CV was performed for 3, 5, 7, 10, and 15 cycles to achieve an optimal thickness and stable polymer layer. The evaluation was based on changes in peak current values after template removal and EP. Beyond ten cycles, the peak current response began to decline (Figure S1B). Consequently, 10 cycles were identified as the most reliable and efficient condition for EP.

**3.3.3. Removal Solution and Time.** Template removal is a pivotal step in MIP preparation, as it ensures the formation of specific recognition sites. To determine the most suitable solvent, several candidates were tested, including acetone, acetonitrile, ethanol, methanol, glacial acetic acid (17.5 M), acetic acid (10 M), as well as binary mixtures of ethanol–acetonitrile (1:1, v/v) and methanol–acetonitrile (1:1, v/v). Among these, the ethanol–acetonitrile mixture exhibited the most excellent efficiency for template extraction, as reflected by the highest  $\Delta I_1$  values (Figure S1C).

Following solvent optimization, the influence of removal time was evaluated using a thermoshaker over 3 to 15 min. The results indicated that  $\Delta I_1$  peaked at 10 min (Figure S1D). Extending the removal period beyond this point resulted in reduced performance, likely due to polymer chain degradation, partial pore collapse, and undesired recross-linking within the matrix. Hence, a removal time of 10 min was selected as the optimal condition for subsequent analyses.

**3.3.4. Rebinding Time.** The rebinding step is a critical parameter that influences both the efficiency and the duration of the analysis. To determine the optimal rebinding time for stable, effective binding, variations in peak current ( $\Delta I_2$ ) before and after rebinding were examined. For this purpose, the MIP-based sensor prepared via EP was immersed in a  $5 \times 10^{-11}$  M IVM solution and tested at different durations (3, 5, 7, 10, 12, 15, and 20 min) using a thermoshaker (650 rpm, 25 °C). The results showed that  $\Delta I_2$  reached its maximum value at 15 min. Therefore, a 15 min rebinding time was selected as the optimal condition for the MAA-IVM@MIP/GCE sensor, as illustrated in Figure S1E.

### 3.4. Analytical Performance of the MIP and NIP Sensors

The electrochemical properties and analytical capability of the MAA-IVM@MIP/GCE sensor were investigated using the  $[\text{Fe}(\text{CN})_6]^{3-/-4-}$  redox probe. DPV was applied as an indirect detection strategy for IVM under optimized conditions. In this method, instead of directly measuring the drug, changes in the probe's electrochemical signal were monitored. When IVM molecules occupy the imprinted cavities on the MIP surface, electron transfer at the electrode interface is partially blocked, leading to a reduction in the redox peak current. This decrease, denoted as  $\Delta I_2$ , increases proportionally with the concentration of IVM.

A calibration curve constructed by plotting  $\Delta I_2$  versus IVM concentration ( $1.0 \times 10^{-12}$  –  $1.0 \times 10^{-11}$  M) exhibited

excellent linearity (Figure 3A). The regression equation was  $\Delta I_2 (\mu\text{A}) = 4.54 \times 10^{12} + 57.39$ , with a correlation coefficient (R) of 0.9918 (Table 1). From the slope of this curve and the

**Table 1. Regression Parameters for the Determination of IVM Using the MAA-IVM@MIP/GCE Sensor**

	Standard solution	Commercial serum sample
Linearity range (M)	$1.0 \times 10^{-12}$ to $1.0 \times 10^{-11}$	$1.0 \times 10^{-12}$ to $1.0 \times 10^{-11}$
Slope ( $\mu\text{A M}^{-1}$ )	$4.54 \times 10^{12}$	$7.81 \times 10^{12}$
SE of slope	$1.79 \times 10^{11}$	$1.18 \times 10^{12}$
Intercept ( $\mu\text{A}$ )	57.39	37.36
SE of intercept	1.68	7.22
Correlation coefficient (R)	0.9918	0.9900
LOD (M)	$2.91 \times 10^{-13}$	$2.91 \times 10^{-13}$
LOQ (M)	$9.71 \times 10^{-13}$	$9.71 \times 10^{-13}$
Repeatability of response (RSD%) <sup>a</sup>	0.79	0.87
Reproducibility of response (RSD%) <sup>a</sup>	1.93	1.67

<sup>a</sup>Each value is the mean of five experiments.

standard deviation of the blank response, the limit of detection (LOD) and limit of quantification (LOQ) were calculated as  $2.91 \times 10^{-13}$  and  $9.71 \times 10^{-13}$  M, respectively.

$$\text{LOD} = 3\sigma/\text{slope} \quad (2)$$

$$\text{LOQ} = 10\sigma/\text{slope} \quad (3)$$

where  $\sigma$  = standard deviation.<sup>29</sup>

Further comparison between MIP- and NIP-modified electrodes confirmed the system's selective recognition capability. The MIP electrode displayed an apparent, concentration-dependent suppression of  $\Delta I_2$  (red curve). In contrast, the NIP electrode showed only negligible variations (blue curve), consistent with the absence of specific binding sites. These observations verify that the MAA-IVM@MIP/GCE sensor provides exceptional sensitivity and selectivity for ultratrace IVM detection.

### 3.5. Biological and Pharmaceutical Product Determination

The MAA-IVM@MIP/GCE sensor was successfully applied for the reliable detection of IVM in both pharmaceutical tablet formulations and spiked human serum samples. To assess its analytical performance, different concentrations of IVM ( $1.0 \times 10^{-12}$  –  $1.0 \times 10^{-11}$  M) were tested in serum. Within this range, the peak current responses exhibited a well-defined linear relationship (Figure 3C), expressed by the regression equation  $\Delta I_2 (\mu\text{A}) = 7.81 \times 10^{12} + 37.36$  with an excellent correlation coefficient (R = 0.9900). The sensor demonstrated remarkable sensitivity, with LOD and LOQ values of  $2.91 \times 10^{-13}$  M and  $9.71 \times 10^{-13}$  M, respectively, calculated from the regression data (Table 1).

Comparative calibration plots of MIP- and NIP-based electrodes in both standard solutions and spiked serum samples (Figures 3A and 3C) further confirmed the sensor's selectivity. The MIP sensor exhibited a clear, concentration-dependent increase in  $\Delta I_2$ , whereas the NIP sensor response remained nearly unchanged. This contrast highlights the superior binding affinity and recognition capability of the imprinted electrode toward IVM. Figures 3A and 3B, respectively, display the resultant DPVs that were created by

rebinding different concentrations of the linear curve in standard solution and serum samples.

To validate the sensor's practical applicability, recovery experiments were performed (Table 2). The recovery values

**Table 2. Recovery Results of IVM Determined with the MAA-IVM@MIP/GCE Sensor in Tablet and Serum Samples**

	Pharmaceutical tablet form	Commercial serum sample
Label amount (mg)	3.00	
Found amount (mg) <sup>a</sup>	2.90	
RSD% <sup>a</sup>	4.65	
Bias%	-3.20	
Spiked amount (mg)	10.00	10.00
Found amount (mg) <sup>a</sup>	9.99	10.29
Average recovery (%)	99.94	102.93
RSD% of recovery <sup>a</sup>	4.20	2.97
Bias%	-0.06	2.93

<sup>a</sup>Each value is the mean of five experiments.

obtained for tablet and serum samples were 99.94% and 102.93%, respectively, confirming the high accuracy and reliability of the method for real sample analysis.

### 3.6. Tap Water Recovery Study

To assess the sensor's practical effectiveness and analytical robustness, recovery experiments were performed using tap water as an environmental test matrix. Assessing tap water illustrates the dependability of the sensor under actual environmental and ecological circumstances, thereby endorsing its viability for sustainable monitoring purposes. Recovery values ranging from 98.07% to 103.04% were obtained, demonstrating the accuracy and reliability of the proposed method for environmental analysis (Table 3).

**Table 3. Recovery Results of IVM in Tap Water Samples with the MAA-IVM@MIP/GCE Sensor**

Sample Concentration (M)	$1 \times 10^{-12}$	$2.5 \times 10^{-12}$	$5 \times 10^{-12}$
Spiked Amount (M)	$6.5 \times 10^{-12}$	$5 \times 10^{-12}$	$2.5 \times 10^{-12}$
Found Amount (M)	$7.71 \times 10^{-12}$	$7.73 \times 10^{-12}$	$7.36 \times 10^{-12}$
Average recovery (%) <sup>a</sup>	102.80	103.04	98.07
RSD% of recovery	3.96	5.79	0.31
Bias%	2.80	3.04	-1.93

<sup>a</sup>Each value is the mean of five experiments.

### 3.7. Soil Extraction Recovery Study

To further evaluate the sensor's performance in complex environmental systems, recovery experiments were carried out using soil samples. As a heterogeneous and demanding matrix, soil offers a realistic measure of the method's resilience and selectivity. The recovery results showed that the sensor can accurately quantify the analyte even in solid samples with significant matrix interference. Overall, these findings confirm that the developed method is reliable, environmentally compatible, and effective for sustainable monitoring of contaminated soil environments (Table 4).

**Table 4. Recovery Results of IVM in Soil Samples with the MAA-IVM@MIP/GCE Sensor**

Sample Concentration (M)	$1 \times 10^{-12}$	$2.5 \times 10^{-12}$	$5 \times 10^{-12}$
Spiked Amount (M)	$6.5 \times 10^{-12}$	$5 \times 10^{-12}$	$2.5 \times 10^{-12}$
Found Amount (M)	$7.47 \times 10^{-12}$	$7.41 \times 10^{-12}$	$7.24 \times 10^{-12}$
Average recovery (%) <sup>a</sup>	99.67	98.78	96.56
RSD% of recovery	3.70	1.67	1.35
Bias%	-0.33	-1.22	-3.44

<sup>a</sup>Each value is the mean of five experiments.

### 3.8. Imprinting Factor/Model Drugs

A primary aim of molecularly imprinted polymers is to achieve high selectivity for the target molecule, particularly in the presence of structurally similar analogs. To investigate this property, a selectivity study was conducted using competitor compounds with comparable chemical structures and properties.

In this study, the model drugs were selected based on their similarity in carbon number and molecular weight to IVM, since compounds with identical functional groups were not readily available. The MAA-IVM@MIP/GCE sensor's selective binding to IVM was assessed against model drugs, including paclitaxel and ritonavir (Table 5). The results

**Table 5. Evaluation of the MAA-IVM@MIP/GCE Sensor's Selectivity for IVM against Compounds of Comparable Size**

	MAA-IVM@MIP/GCE		NIP		$k'_{(MIP/NIP)}$
	$\Delta I_2/\mu A$	$k_{(MIP)}$	$\Delta I_2/\mu A$	$k_{(NIP)}$	
Ivermectin	81.39		58.06		
Paclitaxel	18.07	4.50	28.17	2.06	2.19
Ritonavir	8.96	9.09	12.95	4.48	2.03

showed that the MIP sensor had a markedly higher affinity for IVM, demonstrating its excellent selectivity. This indicates that IVM interacts specifically with the recognition cavities formed during the imprinting process, which are precisely tailored to its size and shape. In contrast, the NIP sensor displayed a weak response, likely due to nonspecific interactions between the functional monomers and the model drugs, as it lacks the specific binding sites for IVM (Figure S2).

### 3.9. Interference

IVM was tested at a concentration of  $5 \times 10^{-12}$  M in the presence of possible interfering species, including  $K^+$ ,  $NO_3^-$ ,  $Na^+$ ,  $SO_4^{2-}$ ,  $Mg^{2+}$ ,  $Cl^-$ , dopamine (DOP), paracetamol (PAR), uric acid (UA), and ascorbic acid (AA). The results indicated that none of these interferents caused a significant change in the IVM peak current response. Recovery values ranged from 97.03% to 104.69% (Table 6). These findings confirm that the analytical efficiency of the MAA-IVM@MIP/GCE sensor remains stable even in the presence of interfering compounds (Figure S3).

### 3.10. Binding Affinity and Isotherm Analysis

In order to assess the originality and efficacy of the proposed MIP sensor, an investigation into the binding affinity was conducted utilizing the Langmuir adsorption model. This model is particularly pertinent for MIP-based electrochemical sensors, as it presumes that the target analyte, IVM, attaches to a limited number of uniform, independent sites within the polymer matrix. The binding constant ( $K_b$ , frequently denoted

**Table 6. Influence of Various Interfering Substances on IVM Detection**

Interference	Recovery of IVM (%)	RSD (%)
KNO <sub>3</sub>	100.11	3.89
MgCl <sub>2</sub>	104.69	1.13
Na <sub>2</sub> SO <sub>4</sub>	101.99	3.18
Dopamine	103.92	4.85
Paracetamol	101.58	1.08
Uric acid	97.03	3.93
Ascorbic acid	98.55	2.01

as  $K_L$ ) and the maximum binding capacity ( $I_{\max}$ ) were ascertained by linearizing the Langmuir equation in the following manner:

$$\frac{1}{I} = \frac{1}{I_{\max}} + \frac{1}{I_{\max}K_a} \cdot \frac{1}{C} \quad (4)$$

$$I_{\max} = \frac{1}{b} \quad (5)$$

$$K_a = \frac{1}{m \cdot I_{\max}} \quad (6)$$

$$K_d = \frac{1}{K_a} \quad (7)$$

By plotting  $1/I$  against  $1/C$ , a linear regression was achieved ( $R^2 \approx 0.996$ ) with a slope ( $m$ ) of  $6.39 \times 10^{-15}$  and a  $Y$ -intercept ( $b$ ) of 0.0103. The remarkable linearity confirms that the MIP-IVM interaction adheres to the Langmuir adsorption model. The saturation-type behavior observed experimentally ( $I_{\max} \approx 97.1$ ) aligns with monolayer adsorption on uniform binding sites. The LOD for an electrochemical sensor is fundamentally associated with the binding constant ( $K_a$ ). A high  $K_a$  value ( $1.61 \times 10^{12} \text{ M}^{-1}$ ) signifies an extraordinarily strong affinity between the IVM molecules and the imprinted cavities. This robust chemical attraction guarantees that even at very low concentrations, the analyte is efficiently captured and retained at the electrode surface, leading to a sensitive signal response and a low LOD. Additionally, the dissociation constant ( $K_d$ ) was determined to be  $6.2 \times 10^{-13} \text{ M}$ . This subpicomolar  $K_d$  value underscores the high thermodynamic stability of the MIP-IVM complex, further validating the sensor's exceptional sensitivity and selectivity in comparison to traditional nonimprinted polymers.<sup>30,31</sup>

### 3.11. Comparison of Selected Analytical Methods

A thorough review of the literature revealed numerous effective strategies for IVM detection. While spectroscopic and chromatographic techniques deliver reliable results, they often require complex sample preparation, expensive instrumentation, and skilled personnel. By contrast, electrochemical methods are simpler and faster but sometimes lack sufficient selectivity for detecting drugs in biological samples. The sensor developed in this study overcomes these limitations, offering enhanced selectivity, rapid analysis, cost-effectiveness, and environmental friendliness. Application to real samples demonstrated that the MAA-IVM@MIP/GCE sensor provides accurate, fast, and user-friendly detection. A summary of various IVM determination methods is provided in Table 7.

### 3.12. Computational Studies

DFT calculations were performed to determine the optimum template:monomer ratio and to elucidate possible template-monomer interactions. The structures of the template-monomer complexes at ratios ranging from 1:1 to 1:5 are presented in Figure 4.


The binding energies of the complexes at ratios ranging from 1:1 to 1:5 were also calculated, and the results are summarized in Table S1. To investigate the effect of template:monomer ratio on binding affinity, both computational and experimental approaches were employed. The results reveal a nonlinear trend, with binding energy increasing up to the 1:3 ratio, plateauing at 1:4, and then decreasing at 1:5. The decline observed at 1:5 suggests that excessive monomer concentration may lead to nonspecific interactions or steric hindrance, reducing the effective binding strength.

In MIP-based sensor design, the strength of interaction between the template molecule and the polymer matrix is a key determinant of sensor performance. While sufficient binding strength is necessary to ensure selective molecular recognition, excessively strong interactions may hinder template removal and reduce the sensor's reusability. Conversely, very weak binding may result in poor selectivity and low signal response. Therefore, the goal is not to achieve the highest possible binding energy, but rather to identify a balance that supports both recognition and regeneration.

Among the tested ratios, the 1:1 complex exhibited a binding energy of  $-57.12 \text{ kJ/mol}$ , which reflects a moderate and well-defined interaction profile. This value is sufficiently strong to support selective recognition, while remaining low enough to allow efficient template removal and sensor regeneration. Importantly, experimental studies confirmed that the 1:1 ratio

**Table 7. Comparative Overview of Analytical Techniques for IVM Determination**

Methodology	Linear range	LOD	Sample	Recovery	Ref
TLC	1–50 ng/g	$\leq 40 \text{ ng/g}$	cattle dung		6
Amperometric flow-injection	0.60–50 $\mu\text{mol/L}$	0.30 $\mu\text{mol/L}$	pharmaceutical drugs and urine samples	95–97%	32
ESI LC-MS/MS	1.0–100.0 ng/mL	10 ng/mL	raw milk	85 to 105%	33
LC-MS/MS	1.5–500 $\mu\text{g/L}$	0.1 $\mu\text{g/kg}$	feces, soil, and sewage	$92.27 \pm 12.01\%$	34
HPLC	2 ppb to 2 ppm	1 ppb	bovine feces	84%	35
HPLC-fluorescence	0.025–5 ng/mL	8.33 pg/mL	camel plasma samples	>70%	36
ic-ELISA and lateral-flow immunochromatographic assay strip	0.09–0.770 ng/mL	0.09 ng/mL	raw milk samples	94% to 112% and for the ICA strip from 110% to 125%.	37
DPV	$9.99 \times 10^{-7}$ to $1.1 \times 10^{-4} \text{ M}$	$2.66 \times 10^{-6} \text{ M}$	urine sample	100.953% to 105.2713	38
MAA-IVM@MIP/GCE	$1 \times 10^{-12}$ to $1 \times 10^{-11} \text{ M}$	$4.31 \times 10^{-12} \text{ M}$	pharmaceutical drugs, tap water, soil	99.94% and 102.93%	This study



**Figure 4.** Geometry-optimized structures of the complexes at template:monomer ratios ranging from 1:1 to 1:5.

yielded the best sensor performance in terms of selectivity, signal intensity, and reproducibility. The agreement between computational predictions and experimental outcomes reinforces the validity of the 1:1 ratio as the optimal formulation for this system.


In addition, for the 1:1 template:monomer, which was also experimentally identified as the optimum ratio, the specific template-monomer interactions are illustrated in Figure S4. The results indicate that, in the template-monomer (1:1) complex, the template-monomer interactions are mediated

through hydrogen bonding. In these interactions, the template molecule, ivermectin, acts as both a hydrogen bond donor and a hydrogen bond donor.

### 3.13. Green Profile of the Proposed Sensor

A sensor's green profile highlights its commitment to environmental sustainability through eco-friendly materials, energy-efficient manufacturing, and minimal waste. This strategy aligns with green chemistry principles by conserving resources, reducing emissions, and using reusable parts to

foster environmental responsibility.<sup>19</sup> To evaluate the study's green profile, three metrics were utilized: BAGI, AGREE, and AGREEMIP. The BAGI tool assesses the relevance of the developed method. The scale ranges from 25 to 60, with scores of 60 or above recommended. In this study, a score of 75 was obtained (Figure 5A). This result confirms that the method is



**Figure 5.** Pictograms for (A) BAGI, (B) AGREE, and (C) AGREEMIP.

effective.<sup>39</sup> Furthermore, reusing the sensor reduces environmental impact and supports resource conservation, in line with GAC's sustainability goals for sensitive and selective IVM detection. The AGREE metric evaluates the method against the 12 principles of GAC on a scale from 0 to 1, where higher scores indicate greater "greenness." The investigation achieved a score of 0.64 with this tool (Figure 5B), highlighting advantages like reduced energy consumption and fewer hazardous chemicals. Nonetheless, the method has certain limitations, such as lacking in situ analysis, incomplete automation, and not accounting for renewable resources.<sup>20</sup> For evaluating the synthesis of the Molecularly Imprinted Polymer (MIP), the AGREEMIP tool was employed. It considers energy needs, reagent environmental impact, and other synthesis factors across 12 criteria to generate a weighted score. In this study, the AGREEMIP score was 0.86, categorizing the sensor as "highly green" (Figure 5C). Therefore, AGREEMIP is a more effective GAC method than other approaches for specifically evaluating the environmental impact of MIP synthesis processes.<sup>40</sup>

#### 4. CONCLUSIONS

This study's electrochemical sensor represents a notable breakthrough in the field, employing molecular imprinting as an effective technique for precise chemical detection. The research primarily aimed to develop a MAA-IVM@MIP/GCE sensor using electropolymerization. The sensor demonstrated

accurate recognition and binding to IVM molecules, offering notable benefits such as higher sensitivity and cost efficiency. Validation results, including recovery rates and RSD%, verified the sensor's reliability and practical application. As the first MIP-based IVM sensor, the MAA-IVM@MIP/GCE system holds considerable promise for future applications, particularly in terms of portability, miniaturization, and rapid analysis. This research presents a versatile method for designing high-performance detection systems that integrate molecular imprinting with electrochemical techniques, effectively enabling accurate detection of IVM in real-world samples. Although the present study focused on steady-state quantitative analysis using DPV, the proposed sensor architecture is compatible with chronoamperometric or flow-based configurations, which could enable real-time monitoring in future developments. Ultimately, this study offers a novel approach to sensitive, rapid IVM detection and highlights the potential of MIP materials for developing future portable, point-of-care (POC), and lab-on-a-chip (LOC) diagnostic devices.

#### ■ ASSOCIATED CONTENT

##### Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acsomega.6c00641>.

Supplementary figures and tables, including optimization parameters, selectivity and interference studies, and theoretical study results (PDF)

#### ■ AUTHOR INFORMATION

##### Corresponding Author

Ayşegül Gölcü – Istanbul Technical University, Faculty of Sciences and Letters, Department of Chemistry, Maslak, 34467 Istanbul, Turkey; [orcid.org/0000-0001-5228-1682](https://orcid.org/0000-0001-5228-1682); Email: [aysgolcu@itu.edu.tr](mailto:aysgolcu@itu.edu.tr)

##### Authors

Zeynep Aydemir – Istanbul Technical University, Faculty of Sciences and Letters, Department of Chemistry, Maslak, 34467 Istanbul, Turkey

Beril S. Kaya – Istanbul Technical University, Faculty of Sciences and Letters, Department of Chemistry, Maslak, 34467 Istanbul, Turkey; Istanbul Health and Technology University, Faculty of Pharmacy, 34445 Istanbul, Turkey

Setareh Dorreh – Istanbul Technical University, Faculty of Sciences and Letters, Department of Chemistry, Maslak, 34467 Istanbul, Turkey

Abdullah Al Faysal – Istanbul Technical University, Faculty of Sciences and Letters, Department of Chemistry, Maslak, 34467 Istanbul, Turkey; Department of Chemistry, Faculty of Science, Atatürk University, Erzurum 25240, Turkey; [orcid.org/0000-0001-6151-074X](https://orcid.org/0000-0001-6151-074X)

Taner Erdoğan – Kocaeli University, Kocaeli Vocational School, Department of Chemistry and Chemical Processing Technologies, 41380 Kocaeli, Turkey

Complete contact information is available at:

<https://pubs.acs.org/10.1021/acsomega.6c00641>

##### Funding

This study was not supported by any funding.

##### Notes

The authors declare no competing financial interest.

## ACKNOWLEDGMENTS

The computational studies reported in this paper were partially performed at TUBITAK ULAKBIM, High Performance and Grid Computing Center (TRUBA resources), and Kocaeli University (Kocaeli University Scientific Project Units, Project No. FBA-2023-3427).

## REFERENCES

- (1) Li, M.; Chen, Z.; Lin, X.; Zhang, X.; Song, Y.; Wen, Y.; Li, J. Engineering of Avermectin Biosynthetic Genes to Improve Production of Ivermectin in *Streptomyces Avermitilis*. *Bioorg. Med. Chem. Lett.* **2008**, *18* (20), 5359–5363.
- (2) Gokbulut, C.; Cirak, V. Y.; Senlik, B.; Aksit, D.; Durmaz, M.; McKellar, Q. A. Comparative Plasma Disposition, Bioavailability and Efficacy of Ivermectin Following Oral and Pour-on Administrations in Horses. *Vet. Parasitol.* **2010**, *170* (1–2), 120–126.
- (3) Heidary, F.; Gharebaghi, R. Ivermectin: A Systematic Review from Antiviral Effects to COVID-19 Complementary Regimen. *J. Antibiot.* Springer Nature September 1, **2020**; 73593–602.
- (4) Rizzo, E. Ivermectin, Antiviral Properties and COVID-19: A Possible New Mechanism of Action. *Naunyn. Schmiedebergs. Arch. Pharmacol.* **2020**, *393* (7), 1153–1156.
- (5) Barati, N.; Motavallihaghi, S.; Nikfar, B.; Chaichian, S.; Momtazi-Borojeni, A. A. Potential Therapeutic Effects of Ivermectin in COVID-19. *Experimental Biology and Medicine*. SAGE Publications Inc. August 1, **2022**; pp 1388–1396, DOI: 10.1177/15353702221099579.
- (6) Floate, K. D.; Taylor, W. G.; Spooner, R. W. Thin-Layer Chromatographic Detection of Ivermectin in Cattle Dung. *J. Chromatogr. B* **1997**; *694*, 246.
- (7) Crooks, S. R. H.; Baxter, A. G.; Traynor, I. M.; Elliott, C. T.; McCaughey, W. J. Detection of Ivermectin Residues in Bovine Liver Using an Enzyme Immunoassay. *Analyst* **1998**, *123*, 355.
- (8) Li, J. S.; Li, W.; Hu, H. B. Immunoaffinity Column Cleanup Procedure for Analysis of Ivermectin in Swine Liver. *J. Chromatogr. B* **1997**; *696*, 166.
- (9) Kennedy, D. G.; Cannavan, A.; Hewitt, S. A.; Rice, D. A.; Blanchflower, W. J. Determination of Ivermectin Residues in the Tissues of Atlantic Salmon (*Salmo Salar*) Using HPLC with Fluorescence Detection. *Food Addit. Contam.* **1993**, *10* (5), 579–584.
- (10) Xie, Z.; Kong, D.; Liu, L.; Song, S.; Kuang, H. Development of Ic-ELISA and Lateral-Flow Immunochromatographic Assay Strip for the Simultaneous Detection of Avermectin and Ivermectin. *Food Agric. Immunol.* **2017**, *28* (3), 439–451.
- (11) Hu, X.; Mutus, B. Electrochemical Detection of Sulfide. *Reviews in Analytical Chemistry*. **2013**, *32*, 247–256.
- (12) Qader, B.; Baron, M.; Hussain, I.; Sevilla, J. M.; Johnson, R. P.; Gonzalez-Rodriguez, J. Electrochemical Determination of Disulfoton Using a Molecularly Imprinted Poly-Phenol Polymer. *Electrochim. Acta* **2019**, *295*, 333–339.
- (13) Zhao, Z.; Teng, Y.; Xu, G.; Zhang, T.; Kan, X. Molecular Imprinted Polymer Based Thermo-Sensitive Electrochemical Sensor for Theophylline Recognition. *Anal. Lett.* **2013**, *46* (14), 2180–2188.
- (14) Luo, J.; Fan, C.; Wang, X.; Liu, R.; Liu, X. A Novel Electrochemical Sensor for Paracetamol Based on Molecularly Imprinted Polymeric Micelles. *Sens. Actuators B Chem.* **2013**, *188*, 909–916.
- (15) Alizadeh, T.; Akbari, A. A Capacitive Biosensor for Ultra-Trace Level Urea Determination Based on Nano-Sized Urea-Imprinted Polymer Receptors Coated on Graphite Electrode Surface. *Biosens. Bioelectron.* **2013**, *43* (1), 321–327.
- (16) Curk, T.; Dobnikar, J.; Frenkel, D. Rational Design of Molecularly Imprinted Polymers. *Soft Matter* **2016**, *12* (1), 35–44.
- (17) Tadi, K. K.; Motghare, R. V. Computational and Experimental Studies on Oxalic Acid Imprinted Polymer. *Computational and Experimental Studies on Oxalic Acid Imprinted Polymer J. Chem. Sci.* **2013**; *125*, DOI: 10.1007/s12039-013-0381-2.
- (18) Qader, B.; Baron, M.; Hussain, I.; Gonzalez-Rodriguez, J. Electrochemical Determination of 2-Isopropoxyphenol in Glassy Carbon and Molecularly Imprinted Poly-Pyrrole Electrodes. *J. Electroanal. Chem.* **2018**, *821*, 16–21.
- (19) Locatelli, M.; Kabir, A.; Perrucci, M.; Ulusoy, S.; Ulusoy, H. I.; Ali, I. Green Profile Tools: Current Status and Future Perspectives. *Advances in Sample Preparation* **2023**, *6*, No. 100068.
- (20) Pena-Pereira, F.; Wojnowski, W.; Tobiszewski, M. AGREE—Analytical GREENess Metric Approach and Software. *Anal. Chem.* **2020**, *92* (14), 10076–10082.
- (21) Lee, C.; Yang, W.; Parr, R. G. Development of the Collesalvetti Correlation-Energy Formula into a Functional of the Electron Density. *Phys. Rev. B* **1988**, *37* (2), 785.
- (22) Becke, A. D. Density-functional Thermochemistry. III. The Role of Exact Exchange. *J. Chem. Phys.* **1993**, *98* (7), 5648–5652.
- (23) Morris, G. M.; Huey, R.; Lindstrom, W.; Sanner, M. F.; Belew, R. K.; Goodsell, D. S.; Olson, A. J. AutoDock4 and AutoDockTools4: Automated Docking with Selective Receptor Flexibility. *J. Comput. Chem.* **2009**, *30* (16), 2785–2791.
- (24) Trott, O.; Olson, A. J. AutoDock Vina: Improving the Speed and Accuracy of Docking with a New Scoring Function, Efficient Optimization, and Multithreading. *J. Comput. Chem.* **2010**, *31* (2), 455–461.
- (25) Frisch, M. J.; Trucks, G. W.; Schlegel, H. B.; Scuseria, G. E.; Robb, M. A.; Cheeseman, J. R.; Scalmani, G.; Barone, V.; Petersson, G. A.; Nakatsuji, H.; Li, X.; Caricato, M.; Marenich, A. V.; Bloino, J.; Janesko, B. G.; Gomperts, R.; Mennucci, B.; Hratchian, H. P.; Ortiz, J. V.; Izmaylov, A. F.; Sonnenberg, J. L.; Williams-Young, D.; Ding, F.; Lipparini, F.; Egidi, F.; Goings, J.; Peng, B.; Petrone, A.; Henderson, T.; Ranasinghe, D.; Zakrzewski, V. G.; Gao, J.; Rega, N.; Zheng, G.; Liang, W.; Hada, M.; Ehara, M.; Toyota, K.; Fukuda, R.; Hasegawa, J.; Ishida, M.; Nakajima, T.; Honda, Y.; Kitao, O.; Nakai, H.; Vreven, T.; Throssell, K.; Montgomery, J. A., Jr.; Peralta, J. E.; Ogliaro, F.; Bearpark, M. J.; Heyd, J. J.; Brothers, E. N.; Kudin, K. N.; Staroverov, V. N.; Keith, T. A.; Kobayashi, R.; Normand, J.; Raghavachari, K.; Rendell, A. P.; Burant, J. C.; Iyengar, S. S.; Tomasi, J.; Cossi, M.; Millam, J. M.; Klene, M.; Adamo, C.; Cammi, R.; Ochterski, J. W.; Martin, R. L.; Morokuma, K.; Farkas, O.; Foresman, J. B.; Fox, D. J. *Gaussian 16 Revision C.01*. 2016.
- (26) Roy Dennington, T. A.; Keith, John, M. M. *GaussView 6*; Semichem Inc: Shawnee Mission, KS, 2016.
- (27) Chang, C.-E.; Gilson, M. K. Tork: Conformational Analysis Method for Molecules and Complexes. *J. Comput. Chem.* **2003**, *24* (16), 1987–1998.
- (28) *BIOVIA Discovery Studio Visualizer*. Dassault Systèmes: San Diego, 2023.
- (29) Ozkan, S. A.; Kauffmann, J.-M.; Zuman, P. *Electroanalysis in Biomedical and Pharmaceutical Sciences*; Springer Berlin Heidelberg: Berlin, Heidelberg, 2015 DOI: 10.1007/978-3-662-47138-8.
- (30) Li, Y.; Luo, L.; Kong, Y.; Li, Y.; Wang, Q.; Wang, M.; Li, Y.; Davenport, A.; Li, B. Recent Advances in Molecularly Imprinted Polymer-Based Electrochemical Sensors. *Biosens. Bioelectron.* **2024**, *249*, No. 116018.
- (31) Sakata, T. Signal Transduction Interfaces for Field-Effect Transistor-Based Biosensors. *Commun. Chem.* **2024**, *7* (1), 35.
- (32) Lourencao, B. C.; Medeiros, R. A.; Thomasi, S. S.; Ferreira, A. G.; Rocha-Filho, R. C.; Fatibello-Filho, O. Amperometric Flow-Injection Determination of the Anthelmintic Drugs Ivermectin and Levamisole Using Electrochemically Pretreated Boron-Doped Diamond Electrodes. *Sens. Actuators B Chem.* **2016**, *222*, 181–189.
- (33) Dahiya, M.; Sen, S.; Lamba, K.; Aggarwal, M.; Khandal, R. K. Quantitative Determination of Ivermectin in Raw Milk Using Positive ESI LC-MS/MS. *Journal of Chemistry*. **2010**, *7*, S267–277.
- (34) Carrillo Heredero, A. M.; Segato, G.; Menotta, S.; Butovskaya, E.; Borra, A.; Bertini, S. Advanced LC-MS/MS Technique for Environmental Ivermectin Detection. *ACS Omega* **2024**, *9* (43), 43649–43657.
- (35) Payne, L. D.; Hicks, M. B.; Wehner, T. A. Determination of Abamectin and/or Ivermectin in Cattle Feces at Low Parts per Billion

Levels Using HPLC with Fluorescence Detection; 1995; Vol. 43.

<https://pubs.acs.org/sharingguidelines>.

(36) Hou, B.; Wang, H.; Jiang, N.; Haosi, B.; Hasi, S. Establishment of the HPLC Fluorescence Detection Method for Plasma Trace Ivermectin and Its Pharmacokinetics in Bactrian Camel. *Vet. Med. Sci.* **2024**, *10* (3), DOI: 10.1002/vms3.1447.

(37) Xie, Z.; Kong, D.; Liu, L.; Song, S.; Kuang, H. Development of Ic-ELISA and Lateral-Flow Immunochromatographic Assay Strip for the Simultaneous Detection of Avermectin and Ivermectin. *Food Agric. Immunol.* **2017**, *28* (3), 439–451.

(38) Koçak, B.; İpek, Y. Electrochemical Detection of Ivermectin Used for the Treatment of COVID-19 with Glutardialdehyde-Modified Glassy Carbon Electrode. *Electrocatalysis* **2022**, *13* (2), 108–115.

(39) Manousi, N.; Wojnowski, W.; Plotka-Wasyłka, J.; Samanidou, V. Blue Applicability Grade Index (BAGI) and Software: A New Tool for the Evaluation of Method Practicality. *Green Chem.* **2023**, *25* (19), 7598–7604.

(40) Marć, M.; Wojnowski, W.; Pena-Pereira, F.; Tobiszewski, M.; Martín-Esteban, A. AGREEMIP: The Analytical Greenness Assessment Tool for Molecularly Imprinted Polymers Synthesis. *ACS Sustain. Chem. Eng.* **2024**, *12* (33), 12516–12524.